

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4224

Name: RITTER 1  
Sterilizer ID: V2178198  
01/24/2024 12:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
2:00 176.1 F 2.9  
4:00 194.9 F 7.0  
6:00 209.7 F 11.8  
8:00 224.8 F 6.0  
10:00 237.2 F 10.9  
12:00 233.1 F 7.6  
14:00 243.6 F 12.1  
16:00 244.7 F 12.8  
18:00 257.9 F 19.4  
20:00 268.8 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.6  
1:00 272.4 F 28.7  
2:00 272.4 F 28.8  
3:00 272.5 F 29.0  
4:00 272.3 F 29.2  
  
Min 271.4 F 28.5  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.6 F 0.0  
5:00 231.2 F 0.0  
10:00 224.5 F 0.0  
15:00 222.2 F 0.0  
20:00 210.1 F 0.0  
25:00 217.2 F 0.0  
30:00 219.8 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 20:58  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:00:26  
CYCLE COMPLETE

01/24/2024 01:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_