
Midmark M11 - v1.0.0
Total Cycles: 5320

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2025 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.0 F 0.0
2:00 125.1 F 1.8
4:00 145.9 F 4.5
6:00 172.2 F 8.3
8:00 194.2 F 12.8
10:00 208.5 F 4.5
12:00 222.7 F 8.8
14:00 234.7 F 13.5
16:00 235.2 F 9.6
18:00 238.1 F 9.3
20:00 246.0 F 13.5
22:00 258.0 F 19.9
24:00 268.9 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.7 F 29.1
2:00 272.2 F 29.1
3:00 272.4 F 28.8
4:00 272.5 F 29.3

Min 271.6 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 192.5 F 0.0
10:00 191.6 F 0.0
15:00 194.9 F 0.0
20:00 198.3 F 0.0
25:00 202.9 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 25:06
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:03:21
CYCLE COMPLETE

02/06/2025 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____