
Midmark M11 - v1.0.0
Total Cycles: 3376

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2023 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.9 F 0.0
2:00 186.1 F 4.6
4:00 205.9 F 9.6
6:00 218.7 F 7.3
8:00 231.7 F 10.3
10:00 231.9 F 7.8
12:00 244.3 F 13.5
14:00 243.9 F 12.4
16:00 256.8 F 19.2
18:00 268.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.1
2:00 271.5 F 28.6
3:00 272.5 F 29.2
4:00 271.7 F 28.5

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 216.0 F 0.0
10:00 233.2 F 0.0
15:00 230.1 F 0.0
20:00 225.5 F 0.0
25:00 222.5 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

06/12/2023 03:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____