
Midmark M11 - v1.0.0
Total Cycles: 1297

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2021 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.7 F 0.0
2:00 149.0 F 1.6
4:00 168.2 F 4.7
6:00 192.6 F 9.3
8:00 218.4 F 6.6
10:00 225.8 F 8.4
12:00 238.0 F 13.5
14:00 239.9 F 10.7
16:00 235.7 F 8.5
18:00 250.7 F 15.4
20:00 262.0 F 22.1
22:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.2 F 28.7
2:00 272.8 F 29.2
3:00 271.6 F 28.6
4:00 272.2 F 28.9

Min 271.4 F 28.4
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 213.7 F 0.0
10:00 192.5 F 0.0
15:00 193.1 F 0.0
20:00 189.7 F 0.0
25:00 187.0 F 0.0
30:00 181.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

10/25/2021 12:35 PM
ACCEPTED: X
REJECTED:
Operator: U01