

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3998

Name: RITTER 1  
Sterilizer ID: V2178198  
11/09/2023 08:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.5 F 0.0  
2:00 72.3 F 0.2  
4:00 84.4 F 1.2  
6:00 117.6 F 3.2  
8:00 142.0 F 5.6  
10:00 164.0 F 8.4  
12:00 179.1 F 11.4  
14:00 203.0 F 3.2  
16:00 208.5 F 5.0  
18:00 217.8 F 8.1  
20:00 227.6 F 11.6  
22:00 226.6 F 6.6  
24:00 238.8 F 11.5  
26:00 235.2 F 8.8  
28:00 247.8 F 14.4  
30:00 258.1 F 20.0  
32:00 267.7 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.7 F 29.3  
2:00 271.7 F 28.6  
3:00 271.8 F 29.1  
4:00 272.4 F 29.1  
  
Min 271.3 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.0 F 0.0  
5:00 202.3 F 0.0  
10:00 204.4 F 0.0  
15:00 207.3 F 0.0  
20:00 210.5 F 0.0  
25:00 210.5 F 0.0  
30:00 217.6 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 33:29  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:11:59  
CYCLE COMPLETE

11/09/2023 10:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_