

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5115

Name: RITTER 1  
Sterilizer ID: V2178198  
11/20/2024 01:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.2 F 0.0  
2:00 167.1 F 0.1  
4:00 195.1 F 0.6  
6:00 211.5 F 3.2  
8:00 224.4 F 7.2  
10:00 232.3 F 11.6  
12:00 233.4 F 7.9  
14:00 244.6 F 13.5  
16:00 243.4 F 11.9  
18:00 241.5 F 10.9  
20:00 254.3 F 17.4  
22:00 265.5 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.4 F 28.9  
2:00 272.2 F 28.6  
3:00 271.7 F 28.1  
4:00 271.4 F 28.2

Min 271.3 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 210.1 F 0.0  
10:00 197.7 F 0.0  
15:00 203.2 F 0.0  
20:00 233.9 F 0.0  
25:00 232.7 F 0.0  
30:00 231.0 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 23:52  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:02:33  
CYCLE COMPLETE

11/20/2024 02:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_