
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006315
9:38 18/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:59
STERILIZING 3:03
136.4 C 222kPa 3:03
136.4 C 225kPa 3:33
136.2 C 224kPa 4:03
136.3 C 224kPa 4:33
136.4 C 225kPa 5:03
136.4 C 224kPa 5:33
136.3 C 224kPa 6:03
136.4 C 224kPa 6:33
136.4 C 224kPa 6:33
VENTING 6:33
AIR DRYING 7:23
CYCLE COMPLETE 9:26

Digital Signature #
507C827D9A27C3B9
