
Midmark M11 - v1.0.5
Total Cycles: 384

Name: RITTER2
Sterilizer ID: V2313228
09/22/2021 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.0 F 0.0
2:00 136.4 F 2.0
4:00 153.6 F 5.3
6:00 181.9 F 9.7
8:00 206.4 F 7.0
10:00 219.6 F 8.6
12:00 233.5 F 12.6
14:00 237.0 F 10.9
16:00 235.5 F 9.3
18:00 250.0 F 16.0
20:00 261.7 F 22.8
22:00 272.3 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.7 F 29.9
2:00 271.5 F 29.7
3:00 272.4 F 29.9
4:00 271.3 F 29.4

Min 271.2 F 29.3
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 238.7 F 0.0
10:00 217.3 F 0.0
15:00 212.0 F 0.0
20:00 207.0 F 0.0
25:00 218.2 F 0.0
30:00 183.4 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:48
CYCLE COMPLETE

09/22/2021 01:00 PM
ACCEPTED: X
REJECTED:
Operator: U01