
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001624
14:25 11/26/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:16
PRESSURIZING 7:04
STERILIZING 8:40
136.0 C 220kPa 8:40
136.5 C 217kPa 9:10
136.2 C 220kPa 9:40
136.1 C 220kPa 10:10
136.2 C 221kPa 10:40
136.3 C 222kPa 11:10
136.2 C 221kPa 11:40
136.1 C 219kPa 12:10
VENTING 12:10
AIR DRYING 13:15
CYCLE COMPLETE 19:26

Digital Signature #
E5D0BC26CA61B5B8
