
Midmark M11 - v1.0.0
Total Cycles: 1875

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2022 04:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.3 F 0.0
2:00 202.6 F 3.4
4:00 208.1 F 8.5
6:00 221.5 F 10.6
8:00 232.3 F 9.2
10:00 237.1 F 9.9
12:00 243.0 F 12.2
14:00 241.2 F 10.8
16:00 254.7 F 17.7
18:00 266.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.8
2:00 271.6 F 28.5
3:00 272.4 F 28.8
4:00 271.4 F 28.3

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 216.0 F 0.0
10:00 225.2 F 0.0
15:00 228.9 F 0.0
20:00 225.5 F 0.0
25:00 223.7 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

04/26/2022 05:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____