
Midmark M11 - v1.0.5
Total Cycles: 921

Name: RITTER2
Sterilizer ID: V2313228
03/29/2022 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.2 F 0.0
2:00 181.8 F 0.3
4:00 199.7 F 3.1
6:00 221.0 F 8.3
8:00 234.2 F 11.6
10:00 236.8 F 10.5
12:00 234.0 F 8.5
14:00 239.9 F 10.2
16:00 248.2 F 14.7
18:00 260.6 F 21.7
20:00 271.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.0 F 29.6
2:00 272.1 F 29.2
3:00 271.5 F 29.1
4:00 272.3 F 29.6

Min 271.4 F 28.8
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 219.2 F 0.0
10:00 237.6 F 0.0
15:00 230.0 F 0.0
20:00 228.6 F 0.0
25:00 230.1 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

03/29/2022 11:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____