
Midmark M11 - v1.0.0
Total Cycles: 4788

Name: RITTER 1
Sterilizer ID: V2178198
08/07/2024 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.9 F 0.0
2:00 138.1 F 0.6
4:00 154.1 F 2.9
6:00 179.2 F 6.2
8:00 205.4 F 10.3
10:00 222.4 F 6.4
12:00 225.1 F 7.4
14:00 233.5 F 11.8
16:00 234.1 F 8.2
18:00 245.2 F 12.6
20:00 243.9 F 12.0
22:00 255.3 F 18.0
24:00 265.8 F 24.2
26:00 271.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.2
2:00 272.6 F 28.8
3:00 271.4 F 28.4
4:00 272.3 F 29.1

Min 271.3 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 235.4 F 0.0
10:00 226.9 F 0.0
15:00 232.9 F 0.0
20:00 229.7 F 0.0
25:00 222.9 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 26:39
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:05:15
CYCLE COMPLETE

08/07/2024 11:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____