
Midmark M11 - v1.0.0
Total Cycles: 4283

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2024 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 75.3 F 0.2
4:00 80.8 F 1.0
6:00 100.4 F 2.8
8:00 132.2 F 5.4
10:00 158.9 F 8.5
12:00 181.0 F 12.2
14:00 203.5 F 3.3
16:00 212.1 F 6.4
18:00 223.2 F 9.9
20:00 230.9 F 10.2
22:00 235.4 F 9.5
24:00 240.0 F 10.4
26:00 244.2 F 12.3
28:00 255.3 F 18.0
30:00 265.0 F 23.9
32:00 271.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.2
3:00 272.5 F 29.0
4:00 271.9 F 28.3

Min 271.3 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 216.5 F 0.0
10:00 215.1 F 0.0
15:00 217.1 F 0.0
20:00 219.4 F 0.0
25:00 221.1 F 0.0
30:00 221.6 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 32:03
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:10:24
CYCLE COMPLETE

02/08/2024 10:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____