

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000789  
7:44 30/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 2:00  
STERILIZING 3:13  
136.5 C 232kPa 3:13  
136.2 C 230kPa 3:43  
136.2 C 231kPa 4:13  
136.3 C 231kPa 4:43  
136.3 C 231kPa 5:13  
136.4 C 232kPa 5:43  
136.3 C 231kPa 6:13  
136.4 C 233kPa 6:43  
136.4 C 233kPa 6:43  
VENTING 6:43  
AIR DRYING 7:34  
CYCLE COMPLETE 7:41  
-----

Digital Signature #  
AFEBCAC257284C  
-----