
Midmark M11 - v1.0.5
Total Cycles: 2527

Name: RITTER2
Sterilizer ID: V2313228
06/19/2023 07:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 75.8 F 0.0
4:00 98.4 F 0.0
6:00 127.2 F 0.0
8:00 135.7 F 2.4
10:00 162.7 F 5.8
12:00 197.8 F 9.3
14:00 221.7 F 13.1
16:00 217.7 F 6.6
18:00 224.8 F 10.6
20:00 230.2 F 8.0
22:00 238.1 F 11.6
24:00 236.9 F 9.4
26:00 249.4 F 15.4
28:00 260.0 F 21.4
30:00 270.0 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 272.3 F 29.5
2:00 272.6 F 29.9
3:00 272.7 F 30.2
4:00 272.0 F 29.8

Min 271.8 F 29.4
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.7 F 0.0
5:00 225.8 F 0.0
10:00 222.6 F 0.0
15:00 234.8 F 0.0
20:00 230.0 F 0.0
25:00 227.4 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 31:38
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:24
CYCLE COMPLETE

06/19/2023 08:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____