
Midmark M11 - v1.0.5
Total Cycles: 3337

Name: RITTER2
Sterilizer ID: V2313228
01/30/2024 01:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.4 F 0.0
2:00 101.4 F 0.0
4:00 119.2 F 0.8
6:00 152.9 F 4.0
8:00 180.2 F 8.3
10:00 202.4 F 12.8
12:00 217.0 F 6.0
14:00 228.7 F 10.4
16:00 231.2 F 7.7
18:00 242.1 F 12.8
20:00 241.2 F 11.3
22:00 253.8 F 17.7
24:00 264.7 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.5
2:00 271.8 F 29.3
3:00 272.3 F 29.3
4:00 271.8 F 29.4

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 233.8 F 0.0
10:00 226.5 F 0.0
15:00 224.4 F 0.0
20:00 223.9 F 0.0
25:00 222.1 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:54
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:04:03
CYCLE COMPLETE

01/30/2024 02:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____