

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 002674  
11:22 11/11/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:29  
PRESSURIZING 2:34  
STERILIZING 3:34  
136.4 C 224kPa 3:34  
135.9 C 222kPa 4:04  
136.4 C 219kPa 4:34  
136.4 C 222kPa 5:04  
136.4 C 224kPa 5:34  
136.4 C 224kPa 6:04  
136.4 C 224kPa 6:34  
136.4 C 222kPa 7:04  
136.4 C 222kPa 7:04  
VENTING 7:04  
AIR DRYING 7:58  
CYCLE COMPLETE 8:04

-----  
Digital Signature #  
FEC616B9A73BAF62  
-----