
Midmark M11 - v1.0.5
Total Cycles: 3971

Name: RITTER2
Sterilizer ID: V2313228
09/24/2024 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.9 F 0.0
2:00 108.6 F 0.0
4:00 118.8 F 0.1
6:00 151.4 F 1.3
8:00 176.3 F 4.8
10:00 194.8 F 8.4
12:00 205.8 F 12.3
14:00 219.9 F 6.5
16:00 228.1 F 11.0
18:00 231.7 F 7.5
20:00 243.0 F 13.2
22:00 242.0 F 11.7
24:00 253.6 F 17.7
26:00 264.2 F 24.0
28:00 271.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 271.6 F 29.4
2:00 272.6 F 29.5
3:00 271.6 F 28.9
4:00 271.7 F 29.5

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 210.9 F 0.0
10:00 223.2 F 0.0
15:00 211.4 F 0.0
20:00 213.6 F 0.0
25:00 228.2 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:01
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:06:31
CYCLE COMPLETE

09/24/2024 02:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____