
Midmark M11 - v1.0.0
Total Cycles: 4940

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2024 08:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 72.1 F 0.0
4:00 93.5 F 0.0
6:00 120.2 F 0.0
8:00 152.6 F 0.1
10:00 181.8 F 0.8
12:00 202.8 F 2.9
14:00 217.3 F 6.0
16:00 225.7 F 9.0
18:00 235.0 F 12.6
20:00 233.5 F 8.0
22:00 243.9 F 12.9
24:00 240.2 F 10.4
26:00 235.6 F 8.1
28:00 250.0 F 14.9
30:00 261.5 F 21.4
32:00 271.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.8 F 28.8
2:00 271.8 F 28.0
3:00 272.1 F 28.5
4:00 272.6 F 28.6

Min 271.6 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 222.3 F 0.0
10:00 184.6 F 0.0
15:00 182.5 F 0.0
20:00 182.7 F 0.0
25:00 236.2 F 0.0
30:00 228.8 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 32:36
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:43
CYCLE COMPLETE

09/26/2024 09:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____