
Midmark M11 - v1.0.0
Total Cycles: 4581

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2024 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.5 F 0.0
2:00 114.1 F 1.1
4:00 137.9 F 3.6
6:00 165.8 F 6.5
8:00 191.5 F 10.0
10:00 208.6 F 7.2
12:00 216.0 F 5.4
14:00 227.2 F 9.1
16:00 235.2 F 13.3
18:00 234.9 F 9.2
20:00 242.9 F 11.4
22:00 243.2 F 12.1
24:00 254.1 F 17.7
26:00 264.2 F 23.6
28:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 271.6 F 28.8
2:00 272.7 F 29.3
3:00 272.2 F 28.9
4:00 271.3 F 28.4

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 210.3 F 0.0
10:00 193.2 F 0.0
15:00 201.1 F 0.0
20:00 210.8 F 0.0
25:00 211.6 F 0.0
30:00 212.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:05
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:06:42
CYCLE COMPLETE

05/24/2024 02:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____