
Midmark M11 - v1.0.0
Total Cycles: 2848

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2023 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.2 F 0.0
2:00 174.4 F 3.5
4:00 184.1 F 7.6
6:00 202.3 F 12.4
8:00 218.3 F 6.4
10:00 231.4 F 11.3
12:00 232.7 F 7.9
14:00 241.7 F 11.9
16:00 244.3 F 12.5
18:00 256.8 F 19.1
20:00 268.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.3
2:00 271.8 F 28.7
3:00 271.5 F 29.0
4:00 272.2 F 29.3

Min 271.2 F 28.5
Max 272.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 208.6 F 0.0
10:00 220.9 F 0.0
15:00 227.8 F 0.0
20:00 229.9 F 0.0
25:00 230.3 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

02/02/2023 12:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____