
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005521
7:51 15/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:47
STERILIZING 3:42
136.0 C 220kPa 3:42
135.9 C 221kPa 4:12
136.0 C 219kPa 4:42
136.4 C 221kPa 5:12
136.4 C 223kPa 5:42
136.4 C 224kPa 6:12
136.3 C 223kPa 6:42
136.3 C 222kPa 7:12
136.3 C 222kPa 7:12
VENTING 7:12
AIR DRYING 8:00
CYCLE COMPLETE 10:01

Digital Signature #
70CBEFBE92F367D8
