

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001484  
10:26 10/07/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:57  
PRESSURIZING 6:41  
STERILIZING 8:06  
136.0 C 220kPa 8:06  
136.3 C 223kPa 8:36  
136.5 C 223kPa 9:06  
136.5 C 223kPa 9:37  
135.9 C 220kPa 10:07  
136.4 C 219kPa 10:37  
136.4 C 224kPa 11:07  
136.3 C 223kPa 11:36  
VENTING 11:37  
AIR DRYING 12:40  
CYCLE COMPLETE 18:42  
-----

Digital Signature #  
1E12DB0644B1B122  
-----