
Midmark M11 - v1.0.5
Total Cycles: 3247

Name: RITTER2
Sterilizer ID: V2313228
12/21/2023 04:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.2 F 0.0
2:00 134.2 F 0.6
4:00 142.2 F 3.3
6:00 171.9 F 6.6
8:00 192.2 F 10.3
10:00 213.8 F 6.0
12:00 217.3 F 8.6
14:00 230.1 F 13.2
16:00 234.0 F 10.0
18:00 237.1 F 9.1
20:00 246.5 F 14.0
22:00 257.7 F 20.1
24:00 267.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.8 F 29.7
2:00 271.5 F 29.1
3:00 271.5 F 29.5
4:00 273.3 F 30.2

Min 270.9 F 28.9
Max 273.3 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 225.0 F 0.0
10:00 230.8 F 0.0
15:00 230.0 F 0.0
20:00 229.6 F 0.0
25:00 228.3 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 25:23
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:59
CYCLE COMPLETE

12/21/2023 05:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____