
Midmark M11 - v1.0.0
Total Cycles: 4356

Name: RITTER 1
Sterilizer ID: V2178198
03/04/2024 11:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.3 F 0.0
2:00 113.9 F 0.6
4:00 116.7 F 2.5
6:00 146.0 F 5.3
8:00 170.7 F 8.5
10:00 190.7 F 12.2
12:00 205.9 F 3.2
14:00 214.4 F 6.7
16:00 225.2 F 10.4
18:00 232.0 F 8.2
20:00 237.9 F 10.6
22:00 236.0 F 8.3
24:00 246.8 F 13.6
26:00 257.1 F 19.1
28:00 266.7 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.5 F 29.0
2:00 271.7 F 28.3
3:00 271.5 F 28.6
4:00 272.4 F 28.9

Min 271.1 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 223.4 F 0.0
10:00 212.4 F 0.0
15:00 218.7 F 0.0
20:00 223.8 F 0.0
25:00 230.6 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:43
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:08:21
CYCLE COMPLETE

03/04/2024 12:49 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____