

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007680  
9:24 16/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:16  
STERILIZING 4:24  
136.3 C 223kPa 4:24  
136.2 C 222kPa 4:54  
136.1 C 218kPa 5:24  
136.3 C 224kPa 5:54  
136.1 C 223kPa 6:24  
136.2 C 220kPa 6:54  
136.3 C 223kPa 7:24  
136.1 C 222kPa 7:54  
136.1 C 222kPa 7:54  
VENTING 7:54  
AIR DRYING 8:47  
CYCLE COMPLETE 8:54

-----  
Digital Signature #  
D1BBC1C573E2690C  
-----