
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001251
15:09 07/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:20
PRESSURIZING 4:20
STERILIZING 5:27
136.0 C 221kPa 5:27
136.2 C 223kPa 5:57
136.4 C 224kPa 6:27
136.4 C 224kPa 6:57
136.2 C 223kPa 7:27
136.4 C 225kPa 7:57
136.4 C 225kPa 8:27
136.2 C 222kPa 8:57
136.2 C 222kPa 8:57
VENTING 8:57
AIR DRYING 9:56
CYCLE COMPLETE 21:04

Digital Signature #
2205E5C0A4BB2775
