
Midmark M11 - v1.0.0
Total Cycles: 697

Name: MIDMARK
Sterilizer ID: V2178198
02/24/2021 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.8 F 0.0
2:00 186.3 F 1.6
4:00 166.7 F 5.9
6:00 196.6 F 11.1
8:00 221.3 F 4.9
10:00 228.6 F 10.7
12:00 233.7 F 7.7
14:00 241.4 F 10.9
16:00 246.4 F 13.2
18:00 259.5 F 20.3
20:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.0 F 28.4
2:00 271.8 F 28.9
3:00 271.9 F 28.6
4:00 272.7 F 29.0

Min 270.9 F 28.0
Max 273.2 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 240.1 F 0.0
10:00 221.8 F 0.0
Dry Cycle Aborted

FILLING: 0:58
HEATING: 20:36
STERILIZING: 4:00
VENTING: 2:55
DRYING: 12:35
TOTAL CYCLE: 00:41:04

02/24/2021 11:53 AM
ACCEPTED: X
REJECTED:
Operator: U01