
Midmark M11 - v1.0.0
Total Cycles: 4838

Name: RITTER 1
Sterilizer ID: V2178198
08/21/2024 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.4 F 0.0
2:00 185.5 F 1.0
4:00 188.5 F 4.8
6:00 211.4 F 9.8
8:00 222.3 F 7.6
10:00 230.1 F 8.5
12:00 239.7 F 13.5
14:00 239.0 F 10.7
16:00 237.2 F 8.7
18:00 250.1 F 15.3
20:00 261.5 F 21.7
22:00 272.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.8
2:00 271.7 F 28.2
3:00 271.9 F 28.7
4:00 272.4 F 28.7

Min 271.4 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 219.1 F 0.0
10:00 220.8 F 0.0
15:00 229.8 F 0.0
20:00 233.5 F 0.0
25:00 226.2 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:21
CYCLE COMPLETE

08/21/2024 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____