
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001076
15:29 13/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:20
STERILIZING 4:22
136.0 C 220kPa 4:22
136.4 C 225kPa 4:52
136.2 C 223kPa 5:22
136.3 C 224kPa 5:52
136.3 C 223kPa 6:22
136.4 C 224kPa 6:52
136.3 C 224kPa 7:22
136.3 C 222kPa 7:52
136.3 C 222kPa 7:52
VENTING 7:52
AIR DRYING 8:45
CYCLE COMPLETE 23:45

Digital Signature #
DC8F7CC469C189DF
