

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3171

Name: RITTER2  
Sterilizer ID: V2313228  
11/30/2023 10:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.5 F 0.0  
2:00 70.8 F 0.0  
4:00 74.1 F 0.2  
6:00 97.5 F 0.8  
8:00 132.0 F 3.0  
10:00 163.4 F 5.8  
12:00 191.2 F 9.5  
14:00 203.7 F 13.4  
16:00 214.9 F 5.9  
18:00 223.9 F 9.8  
20:00 232.7 F 10.4  
22:00 236.6 F 10.8  
24:00 234.6 F 8.6  
26:00 248.0 F 15.0  
28:00 259.0 F 21.1  
30:00 269.2 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.7 F 29.9  
2:00 271.4 F 29.1  
3:00 272.4 F 30.1  
4:00 272.0 F 29.4  
  
Min 271.3 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.3 F 0.0  
5:00 200.2 F 0.0  
10:00 194.2 F 0.0  
15:00 192.4 F 0.0  
20:00 187.9 F 0.0  
25:00 184.4 F 0.0  
30:00 188.9 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:07  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:09:15  
CYCLE COMPLETE

11/30/2023 11:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_