
Midmark M11 - v1.0.0
Total Cycles: 2420

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2022 12:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.0 F 0.0
2:00 181.5 F 4.3
4:00 196.5 F 8.7
6:00 212.3 F 8.0
8:00 227.4 F 8.1
10:00 240.5 F 13.4
12:00 238.7 F 10.7
14:00 235.8 F 8.8
16:00 251.1 F 15.8
18:00 263.1 F 22.8
20:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.3 F 28.9
2:00 271.4 F 28.8
3:00 271.8 F 28.7
4:00 271.6 F 28.9

Min 271.0 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 227.6 F 0.0
10:00 230.8 F 0.0
15:00 229.4 F 0.0
20:00 230.1 F 0.0
25:00 226.6 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

10/12/2022 01:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____