
Midmark M11 - v1.0.0
Total Cycles: 3047

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2023 02:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 155.5 F 2.0
4:00 171.3 F 5.6
6:00 199.9 F 10.1
8:00 219.4 F 5.5
10:00 228.5 F 9.3
12:00 238.0 F 9.8
14:00 241.2 F 11.5
16:00 238.6 F 9.7
18:00 252.3 F 16.4
20:00 263.9 F 23.2
22:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.3 F 28.8
2:00 271.3 F 28.3
3:00 272.4 F 28.9
4:00 271.6 F 28.3

Min 271.2 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 213.6 F 0.0
10:00 204.8 F 0.0
15:00 207.7 F 0.0
20:00 230.1 F 0.0
25:00 229.9 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

03/28/2023 03:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____