
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001552
15:41 30/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:44
STERILIZING 3:43
136.4 C 222kPa 3:43
136.4 C 225kPa 4:13
136.4 C 224kPa 4:43
136.4 C 222kPa 5:13
136.4 C 225kPa 5:43
136.3 C 224kPa 6:13
136.4 C 225kPa 6:43
136.3 C 224kPa 7:13
136.3 C 224kPa 7:13
VENTING 7:13
AIR DRYING 8:04
CYCLE COMPLETE 8:17

Digital Signature #
2D6005724558CB7E
