

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2355

Name: RITTER2  
Sterilizer ID: V2313228  
05/11/2023 11:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.8 F 0.0  
2:00 174.2 F 2.7  
4:00 178.1 F 6.6  
6:00 199.9 F 11.7  
8:00 218.2 F 6.6  
10:00 231.2 F 11.8  
12:00 233.0 F 8.9  
14:00 239.7 F 10.3  
16:00 245.4 F 13.5  
18:00 257.4 F 19.9  
20:00 268.2 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.1 F 29.7  
2:00 272.9 F 29.7  
3:00 272.1 F 29.6  
4:00 272.1 F 29.7  
  
Min 270.8 F 28.8  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.5 F 0.0  
5:00 212.4 F 0.0  
10:00 222.4 F 0.0  
15:00 228.7 F 0.0  
20:00 225.2 F 0.0  
25:00 229.5 F 0.0  
30:00 227.4 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 21:11  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:59:58  
CYCLE COMPLETE

05/11/2023 12:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_