
Midmark M11 - v1.0.5
Total Cycles: 439

Name: RITTER2
Sterilizer ID: V2313228
10/12/2021 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.1 F 0.0
2:00 130.3 F 1.4
4:00 142.3 F 4.3
6:00 171.7 F 8.0
8:00 190.6 F 12.7
10:00 213.6 F 7.0
12:00 226.7 F 11.6
14:00 228.9 F 8.1
16:00 241.5 F 13.1
18:00 241.4 F 12.0
20:00 253.3 F 18.0
22:00 263.9 F 24.3
24:00 271.6 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.4 F 30.5
2:00 271.9 F 30.3
3:00 271.7 F 29.8
4:00 272.4 F 30.2

Min 270.9 F 29.3
Max 273.0 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.9 F 0.0
5:00 227.8 F 0.0
10:00 216.1 F 0.0
15:00 211.8 F 0.0
20:00 189.2 F 0.0
25:00 185.9 F 0.0
30:00 187.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 24:03
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:02:31
CYCLE COMPLETE

10/12/2021 01:16 PM
ACCEPTED: X
REJECTED:
Operator: U01