
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004882
16:11 20/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:24
STERILIZING 3:09
136.0 C 220kPa 3:09
136.3 C 224kPa 3:39
136.3 C 224kPa 4:09
136.4 C 225kPa 4:39
136.3 C 224kPa 5:09
136.4 C 226kPa 5:39
136.4 C 224kPa 6:09
136.4 C 224kPa 6:39
136.4 C 224kPa 6:39
VENTING 6:39
AIR DRYING 7:33
CYCLE COMPLETE 9:16

Digital Signature #
F67919A74F7CBFBE
