
Midmark M11 - v1.0.0
Total Cycles: 2610

Name: RITTER 1
Sterilizer ID: V2178198
11/28/2022 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.3 F 0.0
2:00 203.7 F 4.3
4:00 206.9 F 9.2
6:00 219.6 F 7.4
8:00 231.5 F 10.5
10:00 232.2 F 7.7
12:00 242.0 F 11.1
14:00 246.1 F 13.4
16:00 259.0 F 20.4
18:00 270.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.3
2:00 271.7 F 28.6
3:00 271.7 F 29.1
4:00 272.3 F 29.2

Min 271.2 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 229.7 F 0.0
10:00 228.9 F 0.0
15:00 232.9 F 0.0
20:00 231.3 F 0.0
25:00 229.2 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:49
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

11/28/2022 11:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____