
Midmark M11 - v1.0.0
Total Cycles: 4723

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2024 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.6 F 0.0
2:00 131.6 F 1.6
4:00 146.0 F 4.3
6:00 170.9 F 7.3
8:00 189.3 F 10.5
10:00 208.8 F 3.9
12:00 218.3 F 6.1
14:00 227.6 F 10.0
16:00 232.9 F 9.3
18:00 238.0 F 10.6
20:00 238.5 F 8.5
22:00 247.7 F 14.1
24:00 258.6 F 20.0
26:00 268.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.1
2:00 272.2 F 28.6
3:00 271.4 F 28.3
4:00 272.6 F 29.1

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 219.2 F 0.0
10:00 190.3 F 0.0
15:00 187.7 F 0.0
20:00 201.0 F 0.0
25:00 219.7 F 0.0
30:00 207.3 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 27:18
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:06:07
CYCLE COMPLETE

07/12/2024 12:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____