
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002131
11:35 07/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:46
STERILIZING 3:55
136.0 C 221kPa 3:55
136.3 C 224kPa 4:25
136.1 C 222kPa 4:55
136.4 C 224kPa 5:25
136.2 C 222kPa 5:55
136.4 C 225kPa 6:25
136.4 C 222kPa 6:55
136.4 C 225kPa 7:25
136.4 C 225kPa 7:25
VENTING 7:25
AIR DRYING 8:18
CYCLE COMPLETE 11:55

Digital Signature #
1A215DA49B8A7DF7
