
Midmark M11 - v1.0.5
Total Cycles: 1295

Name: RITTER2
Sterilizer ID: V2313228
07/14/2022 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.3 F 0.0
2:00 171.1 F 0.1
4:00 175.1 F 2.0
6:00 198.3 F 6.7
8:00 215.5 F 12.0
10:00 228.5 F 7.2
12:00 240.2 F 12.4
14:00 237.6 F 10.0
16:00 237.0 F 8.9
18:00 249.0 F 15.0
20:00 260.6 F 21.5
22:00 271.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 29.2
2:00 272.9 F 29.5
3:00 272.2 F 29.8
4:00 271.4 F 29.4

Min 271.2 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 231.5 F 0.0
10:00 226.5 F 0.0
15:00 166.1 F 0.0
20:00 157.0 F 0.0
25:00 160.4 F 0.0
30:00 151.5 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:48
CYCLE COMPLETE

07/14/2022 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____