
Midmark M11 - v1.0.0
Total Cycles: 1951

Name: RITTER 1
Sterilizer ID: V2178198
05/25/2022 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.8 F 0.0
2:00 140.6 F 1.4
4:00 174.5 F 4.2
6:00 205.0 F 7.7
8:00 222.5 F 12.2
10:00 224.3 F 6.2
12:00 234.8 F 10.8
14:00 229.3 F 6.1
16:00 240.5 F 11.4
18:00 235.4 F 8.3
20:00 248.6 F 14.5
22:00 259.6 F 20.6
24:00 270.0 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 271.7 F 28.6
2:00 272.3 F 28.8
3:00 272.7 F 29.0
4:00 272.4 F 29.2

Min 271.4 F 28.6
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.6 F 0.0
5:00 228.5 F 0.0
10:00 205.0 F 0.0
15:00 203.3 F 0.0
20:00 233.5 F 0.0
25:00 215.8 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:39
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:03:40
CYCLE COMPLETE

05/25/2022 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____