
Midmark M11 - v1.0.0
Total Cycles: 4858

Name: RITTER 1
Sterilizer ID: V2178198
08/28/2024 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 143.9 F 0.1
4:00 152.1 F 0.4
6:00 181.7 F 0.9
8:00 200.6 F 4.2
10:00 217.3 F 7.9
12:00 227.9 F 12.2
14:00 229.7 F 7.4
16:00 241.4 F 12.5
18:00 238.4 F 10.0
20:00 237.0 F 8.5
22:00 249.1 F 14.6
24:00 260.4 F 20.9
26:00 270.6 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.7
2:00 271.4 F 28.0
3:00 272.2 F 28.9
4:00 271.9 F 28.3

Min 271.2 F 27.9
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 218.8 F 0.0
10:00 202.4 F 0.0
15:00 221.8 F 0.0
20:00 229.0 F 0.0
25:00 232.5 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 26:52
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:05:16
CYCLE COMPLETE

08/28/2024 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____