
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002517
11:40 19/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:26
STERILIZING 3:33
136.4 C 224kPa 3:33
136.4 C 225kPa 4:03
136.0 C 221kPa 4:33
136.3 C 224kPa 5:03
136.3 C 224kPa 5:33
136.3 C 223kPa 6:03
136.4 C 226kPa 6:33
136.5 C 223kPa 7:03
136.5 C 223kPa 7:03
VENTING 7:03
AIR DRYING 7:55
CYCLE COMPLETE 37:55

Digital Signature #
5AE47D5173B7400A
