
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001461
15:11 18/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:03
STERILIZING 3:07
136.4 C 226kPa 3:07
136.1 C 223kPa 3:37
136.4 C 226kPa 4:07
136.4 C 225kPa 4:37
136.4 C 224kPa 5:07
136.2 C 226kPa 5:37
136.4 C 225kPa 6:07
136.4 C 225kPa 6:37
136.4 C 225kPa 6:37
VENTING 6:37
AIR DRYING 7:27
CYCLE COMPLETE 37:27

Digital Signature #
7BF57464433E8D0A
