
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.3uS / 1.4ppm
CYCLE NUMBER 000055
15:29 05/11/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 6:04
PRESSURIZING 8:13
STERILIZING 10:18
136.0 C 222kPa 10:18
136.3 C 225kPa 10:48
136.2 C 225kPa 11:18
136.3 C 227kPa 11:48
136.5 C 225kPa 12:18
136.4 C 227kPa 12:48
136.3 C 226kPa 13:18
136.3 C 226kPa 13:48
136.3 C 226kPa 13:48
VENTING 13:48
AIR DRYING 15:02
CYCLE COMPLETE 45:02

Digital Signature #
4149B0DE2E84E52D
