
Midmark M11 - v1.0.5
Total Cycles: 3012

Name: RITTER2
Sterilizer ID: V2313228
10/16/2023 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.3 F 0.0
2:00 160.4 F 1.8
4:00 171.8 F 5.5
6:00 199.6 F 10.2
8:00 215.1 F 5.5
10:00 229.2 F 10.3
12:00 231.4 F 7.7
14:00 243.6 F 13.6
16:00 243.3 F 12.5
18:00 256.5 F 19.4
20:00 268.2 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.3 F 29.8
2:00 271.9 F 29.0
3:00 272.2 F 29.8
4:00 272.1 F 29.3

Min 271.2 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 212.2 F 0.0
10:00 205.6 F 0.0
15:00 203.2 F 0.0
20:00 201.0 F 0.0
25:00 199.7 F 0.0
30:00 197.6 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

10/16/2023 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____