

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2295

Name: RITTER 1  
Sterilizer ID: V2178198  
09/07/2022 09:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
2:00 191.9 F 0.8  
4:00 200.7 F 2.8  
6:00 212.4 F 7.9  
8:00 224.8 F 13.0  
10:00 230.9 F 8.0  
12:00 240.5 F 13.4  
14:00 239.0 F 10.7  
16:00 237.3 F 8.6  
18:00 250.4 F 15.3  
20:00 261.6 F 21.8  
22:00 271.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 29.1  
2:00 271.8 F 28.5  
3:00 271.8 F 28.8  
4:00 272.5 F 29.1

Min 271.4 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.7 F 0.0  
5:00 219.8 F 0.0  
10:00 231.0 F 0.0  
15:00 231.7 F 0.0  
20:00 232.0 F 0.0  
25:00 231.8 F 0.0  
30:00 231.2 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 22:38  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:52  
CYCLE COMPLETE

09/07/2022 10:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_