
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001971
9:46 22/08/2024
AIR DRYING ONLY
60 min. MAXIMUM
CYCLE START 0:00
AIR DRYING 0:18
CYCLE COMPLETE 36:30

Digital Signature #
7EE5DFBCE089DB70
