
Midmark M11 - v1.0.5
Total Cycles: 1757

Name: RITTER2
Sterilizer ID: V2313228
11/22/2022 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 214.4 F 0.0
2:00 195.4 F 0.5
4:00 196.6 F 5.0
6:00 218.2 F 10.6
8:00 230.7 F 7.2
10:00 240.9 F 13.5
12:00 241.9 F 12.0
14:00 242.0 F 11.4
16:00 255.6 F 18.6
18:00 267.5 F 26.1
20:00 271.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 273.1 F 29.5
2:00 271.9 F 29.6
3:00 272.4 F 29.4
4:00 272.3 F 29.8

Min 271.4 F 29.0
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 227.7 F 0.0
10:00 228.9 F 0.0
15:00 222.8 F 0.0
20:00 215.6 F 0.0
25:00 225.4 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 20:01
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

11/22/2022 10:02 AM
ACCEPTED: X
REJECTED:
Operator: U02